



Safe-T-Vue® • Product No. 7201-50 and 7201-200
QUALITY ASSURANCE LOT PERFORMANCE REPORT
 Safe-T-Vue® 10 Irreversible Temperature Indicators

Lot # S10 121823

Expiration Date: 2025-06-18

Non - Chemistry

Mark √ for "Accepted" and X for "Not Accepted."

Box	Adh. Label	Foil Lid	Close Function	Fill Level
√ Copy	√ Copy	√ Copy	√ Formed	√ Acceptable
√ Tack	√ Tack	√ Position	√ Locked/Stable	Underfill
√ Clean	√ Color	√ Clean Peel	Hinge	Overfill
√ Box Damage	Tray	Disk	√ Elasticity	Color Label
	√ Edges	√ Appearance		√ Color
	Color	Fit		

Test Protocol – Finished Goods Quality Assurance

Test performed on random sampling of finished product per ANSI ASQZ1.4-2003, General Inspection Level 1, Double Sampling Plan for Attributes with AQL=1%. Electronic thermometers with probes are calibrated annually. Download Validation SOP, from www.zebra.com/safetvue.

Temperature Change and Corresponding Safe-T-Vue® Change	
Product Specification	Safe-T-Vue 6.0±0.3°C, Safe-T-Vue 10±0.3°C
Starting Test Temperatures	1 to 4 °C

Test Parameter	AQL	Sample Size	Accept Number	Test Result	Pass/Fail
Full Color Response Temperature	1%	80	1	0	Pass

Quantity Approved: 22,135

Performance Data: Red Spots - Random 9.1 °C; Red Filling in White Center 9.5 °C; Full Red Color at 9.9 °C.

Quality verification by: G. Thomas Date 2024-01-08

Quality Management approval by: Approved Lot # S10 121823; Rose Dohy Date 2024-01-10