

Safe-T-Vue® • Product No. 7206-50 and 7206-200 QUALITY ASSURANCE LOT PERFORMANCE REPORT

Safe-T-Vue® 6 Irreversible Temperature Indicators

Lot # S6 011823

Expiration Date: 2024-07-18

Non - Chemistry

Mark √ for "Accepted" and X for "Not Accepted."

Вох	Adh. Label	Foil Lid	Close Function	Fill Level
√ Copy	√ Copy	√ Copy	√ Formed	√ Acceptable
√ Tack	√ Tack	√ Position	√ Locked/Stable	Underfill
√ Clean	√ Color	√ Clean Peel	Hinge	Overfill
√ Box Damage	Tray	Disk	√ Elasticity	Color Label
	√ Edges	√ Appearance		√ Color
	Color	√ Fit		 -

Test Protocol - Finished Goods Quality Assurance

Test performed on random sampling of finished product per ANSI ASQZ1.4-2003, General Inspection Level 1, Double Sampling Plan for Attributes with AQL=1%. Electronic thermometers with probes are calibrated annually. Download Validation SOP, from www.zebra.com/safetvue.

Temperature Change and Corresponding Safe-T-Vue® Change				
Product Specification	Safe-T-Vue 6.0±0.3°C, Safe-T-Vue 10±0.3°C			
Starting Test Temperatures	1 to 4 °C			

Test Parameter	AQL	Sample Size	Accept Number	Test Result	Pass/Fail
Full Color Response Temperature	1%	50	0	0	Pass

Quantity Approved: 7,059		

Performance Data: Red Spots - Random 5.0° C; Red Filling in White Center 5.5 ° C; Full Red Color at 6.0° C.

Quality Verification by: <u>Ushomas</u> Date <u>2023-07-18</u>

Quality Management approval by: Approved Lot # <u>SIGO 1823</u>; <u>Rose Day</u> Date <u>2023-07-18</u>

QCLF QAP0102-2 STV QA Lot Performance Report STV Rev. 18

Registry Doc. #2046

Manufactured by: Temptime Corporation, 116 The American Road, Morris Plains, NJ 07950, USA
Tel: +1 (973) 630-6000 Fax: +1 (973) 630-6102 www.zebra.com