

Safe-T-Vue® • Product No. 7201-50 and 7201-200 QUALITY ASSURANCE LOT PERFORMANCE REPORT

Safe-T-Vue® 10 Irreversible Temperature Indicators

Lot # S10 022724

Expiration Date: 2025-08-27

Non - Chemistry

Mark √ for "Accepted" and X for "Not Accepted."

Box	Adh. Label	Foil Lid	Close Function	Fill Level	
√ Copy	√ Copy	√ Copy	√ Formed	√ Acceptable	
√ Tack	√ Tack	√ Position	√ Locked/Stable	Underfill	
√ Clean	√ Color	√ Clean Peel	Hinge	Overfill	
√ Box Damage	Tray Disk		√ Elasticity	Color Label	
	√ Edges	√ Appearance		√ Color	
	√ Color	√ Fit			

Test Protocol – Finished Goods Quality Assurance

Test performed on random sampling of finished product per ANSI ASQZ1.4-2003, General Inspection Level 1, Double Sampling Plan for Attributes with AQL=1%. Electronic thermometers with probes are calibrated annually. Download Validation SOP, from www.zebra.com/safetyue.

Temperature Change and Corresponding Safe-T-Vue® Change			
Product Specification	Safe-T-Vue 6.0±0.3°C, Safe-T-Vue 10±0.3°C		
Starting Test Temperatures	1 to 4 °C		

Test Parameter	AQL	Sample Size	Accept Number	Test Result	Pass/Fail	
Full Color Response Temperature	1%	125	2	0	Pass	

, ,,					
Performance Data	Red Spots - Random	9.4 ° C: Red Filling in White Center	9.6 ° C	Full Red Color at	10000

Quantity Approved: 53.936

Quality Management approval by: Approved Lot # 510 022734; Love Date 2024-04-16

QCLF QAP0102-2 STV QA Lot Performance Report STV Rev. 18 Registry Doc. #2046

Manufactured by: Temptime Corporation, 116 The American Road, Morris Plains, NJ 07950, USA

Tel: +1 (973) 630-6000 Fax: +1 (973) 630-6102 www.zebra.com