



Safe-T-Vue® • Product No. 7201-50 and 7201-200
QUALITY ASSURANCE LOT PERFORMANCE REPORT
 Safe-T-Vue® 10 Irreversible Temperature Indicators

Lot # S10 081722

Expiration Date: 2024-08-17

Non - Chemistry

Mark for "Accepted" and X for "Not Accepted."

Box	Adh. Label	Foil Lid	Close Function	Fill Level
<input checked="" type="checkbox"/> Copy	<input checked="" type="checkbox"/> Copy	<input checked="" type="checkbox"/> Copy	<input checked="" type="checkbox"/> Formed	<input checked="" type="checkbox"/> Acceptable
<input checked="" type="checkbox"/> Tack	<input checked="" type="checkbox"/> Tack	<input checked="" type="checkbox"/> Position	<input checked="" type="checkbox"/> Locked/Stable	<input type="checkbox"/> Underfill
<input checked="" type="checkbox"/> Clean	<input checked="" type="checkbox"/> Color	<input checked="" type="checkbox"/> Clean Peel	Hinge	<input type="checkbox"/> Overfill
<input checked="" type="checkbox"/> Box Damage	Tray	Disk	<input checked="" type="checkbox"/> Elasticity	Color Label
	<input checked="" type="checkbox"/> Edges	<input checked="" type="checkbox"/> Appearance		<input checked="" type="checkbox"/> Color
	<input checked="" type="checkbox"/> Color	<input checked="" type="checkbox"/> Fit		

Test Protocol – Finished Goods Quality Assurance

Test performed on random sampling of finished product per ANSI ASQZ1.4-2003, General Inspection Level 1, Double Sampling Plan for Attributes with AQL=1%. Electronic thermometers with probes are calibrated annually. Download Validation SOP, from www.zebra.com/safetvue.

Temperature Change and Corresponding Safe-T-Vue® Change	
Product Specification	Safe-T-Vue 6.0±0.3°C, Safe-T-Vue 10±0.3°C
Starting Test Temperatures	1 to 4 °C

Test Parameter	AQL	Sample Size	Accept Number	Test Result	Pass/Fail
Full Color Response Temperature	1%	125	2	0	Pass

Quantity Approved: 60,345

Performance Data: Red Spots - Random 9.4 °C; Red Filling in White Center 9.8 °C; Full Red Color at 10.2 °C.

Quality verification by: G. Thomas Date 2022-09-22

Quality Management approval by: Approved Lot # S10 081722; Ravi Dey Date 2022-09-27