

Safe-T-Vue® • Product No. 7206-50 and 7206-200 QUALITY ASSURANCE LOT PERFORMANCE REPORT

Safe-T-Vue® 10 Irreversible Temperature Indicators

Lot # S6 020123

Expiration Date: 2024-08-01

Non - Chemistry

Mark √ for "Accepted" and X for "Not Accepted."

Вох	Adh. Label	Foil Lid	Close Function	Fill Level	
√ Copy	√ Copy	√ Copy	√ Formed	√ Acceptable	
√ Tack	√ Tack	√ Position	√ Locked/Stable	Underfill	
√ Clean	√ Color	√ Clean Peel	Hinge	Overfill	
√ Box Damage	Tray Disk		√ Elasticity	Color Label	
	√ Edges	√ Appearance		√ Color	
	√ Color	√ Fit			

Test Protocol - Finished Goods Quality Assurance

Test performed on random sampling of finished product per ANSI ASQZ1.4-2003, General Inspection Level 1, Double Sampling Plan for Attributes with AQL=1%. Electronic thermometers with probes are calibrated annually. Download Validation SOP, from www.zebra.com/safetvue.

Temperature Change and Corresponding Safe-T-Vue® Change				
Product Specification	Safe-T-Vue 6.0±0.3°C, Safe-T-Vue 10±0.3°C			
Starting Test Temperatures	1 to 4 °C			

Test Parameter	AQL	Sample Size	Accept Number	Test Result	Pass/Fail
Full Color Response Temperature	1%	50	0	0	Pass

Quantity Approved:	5,150				
	Spots - Random <u>5.2° C;</u> R	Red Filling in Wh	ite Center <u>5.6 °</u>	C; Full Red Color a	at <u>5.9° C</u> .
Quality verification by:	Gylomas	Date _	2023-08-11		
Quality Management app	proval by: Approved Lot	5602017	-3: M	4	Date 2023-08-15
OCLE OAPO102-2 STV	ΟΔ Lot Performance R	enort STV Re	v 18	Registry Doc	#2046

Manufactured by: Temptime Corporation, 116 The American Road, Morris Plains, NJ 07950, USA
Tel: +1 (973) 630-6000 Fax: +1 (973) 630-6102 www.zebra.com