

**Safe-T-Vue® • Product No. 7201-50 and 7201-200**  
**QUALITY ASSURANCE LOT PERFORMANCE REPORT**  
 Safe-T-Vue® 10 Irreversible Temperature Indicators

Lot # S10 021521

Expiration Date: 2023-02-15

**Non - Chemistry**

Mark √ for "Accepted" and X for "Not Accepted."

Box	Adh. Label	Foil Lid	Close Function	Fill Level
√ Copy	√ Copy	√ Copy	√ Formed	√ Acceptable
√ Tack	√ Tack	√ Position	√ Locked/Stable	Underfill
√ Clean	√ Color	√ Clean Peel	<b>Hinge</b>	Overfill
√ Box Damage	<b>Tray</b>	<b>Disk</b>	√ Elasticity	<b>Color Label</b>
	√ Edges	√ Appearance		√ Color
	√ Color	√ Fit		

**Test Protocol – Finished Goods Quality Assurance**

Test performed on random sampling of finished product per ANSI ASQZ1.4-2003, General Inspection Level 1, Double Sampling Plan for Attributes with AQL=1%. Electronic thermometers with probes are calibrated annually. Download Validation SOP, from [www.zebra.com/safetvue](http://www.zebra.com/safetvue).

<b>Temperature Change and Corresponding Safe-T-Vue® Change</b>	
Product Specification	Safe-T-Vue 6.0±0.3°C, Safe-T-Vue 10±0.3°C
Starting Test Temperatures	1 to 4 °C

Test Parameter	AQL	Sample Size	Accept Number	Test Result	Pass/Fail
Full Color Response Temperature	1%	125	2	0	PASS

Quantity Approved: 64,144

**Performance Data:** Red Spots - Random 9.3° C; Red Filling in White Center 9.6° C; Full Red Color at 10.0° C.

Quality Control Verification: Maddeline Accardo 2021-03-09  
 Quality manager Date

Quality Assurance: Approved Lot # S10 021521; [Signature] 2021-03-09  
 Quality manager Date