

**Safe-T-Vue® • Product No. 7201-50 and 7201-200**  
**QUALITY ASSURANCE LOT PERFORMANCE REPORT**  
 Safe-T-Vue® 10 Irreversible Temperature Indicators

Lot # **S10 102220**

Expiration Date: **2022-10-22**

**Non - Chemistry**

Mark √ for "Accepted" and X for "Not Accepted."

| Box          | Adh. Label  | Foil Lid     | Close Function  | Fill Level         |
|--------------|-------------|--------------|-----------------|--------------------|
| √ Copy       | √ Copy      | √ Copy       | √ Formed        | √ Acceptable       |
| √ Tack       | √ Tack      | √ Position   | √ Locked/Stable | Underfill          |
| √ Clean      | √ Color     | √ Clean Peel | <b>Hinge</b>    | Overfill           |
| √ Box Damage | <b>Tray</b> | <b>Disk</b>  | √ Elasticity    | <b>Color Label</b> |
|              | √ Edges     | √ Appearance |                 | √ Color            |
|              | √ Color     | √ Fit        |                 |                    |

**Test Protocol – Finished Goods Quality Assurance**

Test performed on random sampling of finished product per ANSI ASQZ1.4-2003, General Inspection Level 1, Double Sampling Plan for Attributes with AQL=1%. Electronic thermometers with probes are calibrated annually. Download Validation SOP, from [www.zebra.com/safetvue](http://www.zebra.com/safetvue).

| <b>Temperature Change and Corresponding Safe-T-Vue® Change</b> |   |
|--|---|
| Product Specification  | Safe-T-Vue 6.0±0.3°C, Safe-T-Vue 10±0.3°C |
| Starting Test Temperatures                                     | 1 to 4 °C                                 |

| Test Parameter                  | AQL | Sample Size | Accept Number | Test Result | Pass/Fail |
|---------------------------------|-----|-------------|---------------|-------------|-----------|
| Full Color Response Temperature | 1%  | 125         | 2             | 1           | Pass      |

Quantity Approved: 50,365

**Performance Data:** Red Spots - Random 9.3 °C; Red Filling in White Center 9.7 °C; Full Red Color at 10.0 °C.

Quality Control Verification: *ncardelino@zebra.com* 2020-12-07  
 Quality manager Date

Quality Assurance: Approved Lot # S10102220; *Allwagner* 2020-12-09  
 Quality manager Date