

Safe-T-Vue® • Product No. 7201-50 and 7201-200 QUALITY ASSURANCE LOT PERFORMANCE REPORT

Safe-T-Vue® 10 Irreversible Temperature Indicators

Lot # S10 022822

Expiration Date: 2024-02-28

Non - Chemistry

Mark √ for "Accepted" and X for "Not Accepted."

Вох		Adh. Label	Foil Lid	Close Function	Fill Level	
	Сору	√ Copy	√ Copy	√ Formed	√ Acceptable	
	Tack	√ Tack	√ Position	√ Locked/Stable	Underfill	
	Clean	√ Color	√ Clean Peel	Hinge	Overfill	
	Box Damage	Tray	Disk	√ Elasticity	Color Label	
		√ Edges	√ Appearance		√ Color	
		√ Color	√ Fit			

Test Protocol – Finished Goods Quality Assurance

Test performed on random sampling of finished product per ANSI ASQZ1.4-2003, General Inspection Level 1, Double Sampling Plan for Attributes with AQL=1%. Electronic thermometers with probes are calibrated annually. Download Validation SOP, from www.zebra.com/safetvue.

Temperature Change	Temperature Change and Corresponding Safe-T-Vue® Change				
Product Specification	Safe-T-Vue 6.0±0.3°C, Safe-T-Vue 10±0.3°C				
Starting Test Temperatures	1 to 4 °C				

Test Parameter	AQL	Sample Size	Accept Number	Test Result	Pass/Fail
Full Color Response Temperature	1%	125	2	0	Pass

Quantity Approved: 62,813		
Performance Data: Red Spots - Random 9.3 ° C;	Red Filling in White Center <u>9.6 ° C</u> ; Full Re	d Color at <u>10.1 ° C</u> .
Quality Control Verification: Quality manager		
Quality Assurance: Approved Lot # <u>SID りぽ</u>	Quality manager	<i>2677-03-18</i> Date

QCLF QAP0102-2 STV QA Lot Performance Report STV Rev. 17 Registry Doc. #2046