ZEBRA | Temptime
Now a part of Zebra Technologies

Safe-T-Vue® • Product No. 7201-50 and 7201-200
QUALITY ASSURANCE LOT PERFORMANCE REPORT
Safe-T-Vue® 10 Irreversible Temperature Indicators

Lot # S10 032221

Expiration Date: **2023-03-22**

Non - Chemistry

Mark √ for "Accepted" and X for "Not Accepted."

Box	Adh. Label	Foil Lid	Close Function	Fill Level
√ Copy	√ Copy	√ Copy	√ Formed	√ Acceptable
√ Tack	√ Tack	√ Position	√ Locked/Stable	Underfill
√ Clean	√ Color	√ Clean Peel	**Hinge**	Overfill
√ Box Damage	**Tray**	**Disk**	√ Elasticity	**Color Label**
	√ Edges	√ Appearance		√ Color
	√ Color	√ Fit		

Test Protocol – Finished Goods Quality Assurance

Test performed on random sampling of finished product per ANSI ASQZ1.4-2003, General Inspection Level 1, Double Sampling Plan for Attributes with AQL=1%. Electronic thermometers with probes are calibrated annually. Download Validation SOP, from www.zebra.com/safetvue.

Temperature Change and Corresponding Safe-T-Vue® Change	
Product Specification	Safe-T-Vue 6.0±0.3°C, Safe-T-Vue 10±0.3°C
Starting Test Temperatures	1 to 4 °C

Test Parameter	AQL	Sample Size	Accept Number	Test Result	Pass/Fail
Full Color Response Temperature	1%	125	2	0	Pass

Quantity Approved: 55233

Performance Data: Red Spots - Random **9.3 ° C**; Red Filling in White Center **9.6 ° C**; Full Red Color at **10.0 ° C**.

Quality Control Verification: Madeline Acevedo — Quality manager; 2021-04-15 — Date

Quality Assurance: Approved Lot # S10 032221 ; [signature] 2021-04-15