

Safe-T-Vue® • Product No. Choose an item.
QUALITY ASSURANCE LOT PERFORMANCE REPORT
 Safe-T-Vue® Choose an item. Irreversible Temperature Indicators

Lot # **S10 031220**

Expiration Date: **2022-03-12**

Non - Chemistry

Mark for "Accepted" and X for "Not Accepted."

Box	Adh. Label	Foil Lid	Close Function	Fill Level
<input checked="" type="checkbox"/> Copy	<input checked="" type="checkbox"/> Copy	<input checked="" type="checkbox"/> Copy	<input checked="" type="checkbox"/> Formed	<input checked="" type="checkbox"/> Acceptable
<input checked="" type="checkbox"/> Tack	<input checked="" type="checkbox"/> Tack	<input checked="" type="checkbox"/> Position	<input checked="" type="checkbox"/> Locked/Stable	<input type="checkbox"/> Underfill
<input checked="" type="checkbox"/> Clean	<input checked="" type="checkbox"/> Color	<input checked="" type="checkbox"/> Clean Peel	Hinge	<input type="checkbox"/> Overfill
<input checked="" type="checkbox"/> Box Damage	Tray	Disk	<input checked="" type="checkbox"/> Elasticity	Color Label
	<input checked="" type="checkbox"/> Edges	<input checked="" type="checkbox"/> Appearance		<input checked="" type="checkbox"/> Color
	<input checked="" type="checkbox"/> Color	<input checked="" type="checkbox"/> Fit		

Test Protocol – Finished Goods Quality Assurance

Test performed on random sampling of finished product per ANSI ASQZ1.4-2003, General Inspection Level 1, Double Sampling Plan for Attributes with AQL=1%. Electronic thermometers with probes are calibrated annually. Download Validation SOP, from www.temptimecorp.com/support.

Temperature Change and Corresponding Safe-T-Vue® Change	
Product Specification	Safe-T-Vue 6.0±0.3°C, Safe-T-Vue 10±0.3°C
Starting Test Temperatures	1 to 4 °C

Test Parameter	AQL	Sample Size	Accept Number	Test Result	Pass/Fail
Full Color Response Temperature	1%	50	0	0	Pass

Quantity Approved: 4,843

Performance Data: Red Spots - Random **9.6 °C**; Red Filling in White Center **9.9 °C**; Full Red Color at **10.1 °C**.

Quality Control Verification: *Madeline Lucero* 2020-05-19
 Quality manager Date

Quality Assurance: Approved Lot # S10031220 *JM Stanton* 2020-05-20
 Quality manager Date