

Safe-T-Vue® • Product No. 7206-50 and 7206-200
QUALITY ASSURANCE LOT PERFORMANCE REPORT
 Safe-T-Vue® 6 Irreversible Temperature Indicators

Lot # S6 060120

Expiration Date: 2022-06-01

Non - Chemistry

Mark ✓ for "Accepted" and X for "Not Accepted."

| Box | Adh. Label | Foil Lid | Close Function | Fill Level |
|--------------|-------------|--------------|-----------------|--------------------|
| ✓ Copy | ✓ Copy | ✓ Copy | ✓ Formed | ✓ Acceptable |
| ✓ Tack | ✓ Tack | ✓ Position | ✓ Locked/Stable | Underfill |
| ✓ Clean | ✓ Color | ✓ Clean Peel | Hinge | Overfill |
| ✓ Box Damage | Tray | Disk | ✓ Elasticity | Color Label |
| | ✓ Edges | ✓ Appearance | | ✓ Color |
| | ✓ Color | ✓ Fit | | |

Test Protocol – Finished Goods Quality Assurance

Test performed on random sampling of finished product per ANSI ASQZ1.4-2003, General Inspection Level 1, Double Sampling Plan for Attributes with AQL=1%. Electronic thermometers with probes are calibrated annually. Download Validation SOP, from www.temptimecorp.com/support.

| Temperature Change and Corresponding Safe-T-Vue® Change | |
|--|---|
| Product Specification | Safe-T-Vue 6.0±0.3°C, Safe-T-Vue 10±0.3°C |
| Starting Test Temperatures | 1 to 4 °C |

| Test Parameter | AQL | Sample Size | Accept Number | Test Result | Pass/Fail |
|---------------------------------|-----|-------------|---------------|-------------|-----------|
| Full Color Response Temperature | 1% | 125 | 2 | 0 | Pass |

Quantity Approved: 51,956

Performance Data: Red Spots - Random 4.9 °C; Red Filling in White Center 5.4 °C; Full Red Color at 5.9 °C.

Quality Control Verification: B DePaema 2020-07-24
 Quality manager Date

Quality Assurance: Approved Lot # S6060120; Allyson 2020-07-24
 Quality manager Date