

Safe-T-Vue® • Product No. 7206-50 and 7206-200
QUALITY ASSURANCE LOT PERFORMANCE REPORT
 Safe-T-Vue® 6 Irreversible Temperature Indicators

Lot # **S6 101119**

Expiration Date: **2021-10-11**

Non - Chemistry

Mark \checkmark for "Accepted" and X for "Not Accepted."

Box	Adh. Label	Foil Lid	Close Function	Fill Level
\checkmark Copy	\checkmark Copy	\checkmark Copy	\checkmark Formed	\checkmark Acceptable
\checkmark Tack	\checkmark Tack	\checkmark Position	\checkmark Locked/Stable	Underfill
\checkmark Clean	\checkmark Color	\checkmark Clean Peel	Hinge	Overfill
\checkmark Box Damage	Tray	Disk	\checkmark Elasticity	Color Label
	\checkmark Edges	\checkmark Appearance		\checkmark Color
	\checkmark Color	\checkmark Fit		

Test Protocol – Finished Goods Quality Assurance

Test performed on random sampling of finished product per ANSI ASQZ1.4-2003, General Inspection Level 1, Double Sampling Plan for Attributes with AQL=1%. Electronic thermometers with probes are calibrated annually. Download Validation SOP, from www.temptimecorp.com/support.

Temperature Change and Corresponding Safe-T-Vue® Change	
Product Specification	Safe-T-Vue 6.0±0.3°C, Safe-T-Vue 10±0.3°C
Starting Test Temperatures	1 to 4 °C

Test Parameter	AQL	Sample Size	Accept Number	Test Result	Pass/Fail
Full Color Response Temperature	1%	80	1	0	Pass

Quantity Approved: 25,684

Performance Data: Red Spots - Random 5.0° C; Red Filling in White Center 5.5° C; Full Red Color at 5.9° C.

Quality Control Verification: *M. Starni* Quality manager 2019-11-11 Date

Quality Assurance: Approved Lot # S6101119; *A. Wagner* Quality manager 2019-11-11 Date